

AUDIOMETERS															
Manufacturer	Frequency Range (Hz)	Attenuation Range (db)	Test Program/Control Options			Test Results Options			Time Per Test (min)	# of Persons Tested Concurrently From One Unit	Is a Booth Required		Audiometer Calibration		# of Validity Checks
			Manual	Automatic	Computer Automated	Printout	Database Stored	Automatic Data Processing			Yes	No	Internal	External	
Beltone Electronics Corp. 773/583-3600	125-8000	-10-110	•						1		•		•		
Maico Diagnostics 888/941-4201	500-8000	-10-100 HL	•	•	•	•	•	•	6	1-8		•		•	
Micro Audiometrics Corp. 828/644-0771	125-8000	-10-100	•	•	•	•	•	•	3-5	1		•		•	
Occupational Health Dynamics 205/980-0180, 888/464-3872	500-8000	0-100	•	•	•	•	•	•	3-5	16		•		•	
Tremetrics Occ. Health Group 800/825-0121	500-8000	-10-100	•	•	•	•	•	•	3-6	8		•	•	•	